



REFERENCE SHEET

REFERENCE MATERIAL

IAEA-313

²²⁶Ra, Th and U IN STREAM SEDIMENT

Date of issue: January 2000[⊕]

Recommended Values
(Based on dry weight)

Reference Date: 30th January 1988

Element	Recommended Value Bq/kg	95% Confidence Interval Bq/kg	N*
²²⁶ Ra	343	307 – 379	24

Element	Recommended Value mg/kg	95% Confidence Interval mg/kg	N*
Th	77.1	74.8 – 79.4	23
U	18.2	17.0 – 19.3	31

* Number of accepted laboratory means which were used to calculate the recommended values and confidence intervals.

⊕ Revision of the original reference sheet dated January 1991

The values listed above were established on the basis of statistically valid results submitted by laboratories which had participated in an international intercomparison exercise organized during 1988. The details concerning the criteria for qualification as a recommended value can be found in the report (IAEA/AL/037) "Report on the Intercomparison Run IAEA-313: ²²⁶Ra, Th and U in Stream Sediment" [1]. This report is available free of charge upon request.

Intended Use

This sample is intended to be used as a reference material for the measurement of ^{226}Ra , Th and U in sediment samples. It can also be used as a quality control material for the assessment of a laboratory's analytical work, for the validation of analytical methods and for quality assurance within a laboratory.

Origin and preparation of the material

The stream sediment sample was collected from the Sibolga area of the west coast of northern Sumatra, Indonesia and was donated by the Indonesian Atomic Energy Commission (IAEC).

The material was air dried and then ground to pass through a 0.18 mm sieve before being homogenized. The material was dispensed into plastic bottles in 50 g units without any further processing. Subsequently, the samples were irradiated to a dose of 2.5×10^4 Gy using a ^{60}Co source to ensure long-term stability of the material by inhibiting microbial action.

Homogeneity

The homogeneity of the bottled material was assessed by the determination of uranium using laser fluorimetry in two 500 mg sub-samples taken from a number of bottles chosen at random from the whole production run. The between and within bottle variability was assessed (on the basis of the F test, at the 95% confidence level) and indicated that the contribution due the heterogeneity of the material was small and therefore the material could be considered to be homogeneous for a sample size at or above 500 mg.

Dry weight determination

All recommended values are expressed on a dry weight basis. Therefore the dry weight must be determined at the time of analysis, using separate sub-samples of at least 500 mg dried to constant weight in a drying oven set to 105 °C. Subsequent weighings should differ by less than 5 mg.

Instructions for use

The recommended sample size for analysis is 500 mg. Analysts are reminded to take appropriate precautions in order to avoid contamination of the material during handling. No special precautions are required for the storage of this material.

Legal disclaimer

The IAEA makes no warranties, expressed or implied, with respect to the data contained in this reference sheet and shall not be liable for any damage that may result from the use of such data.

References

- [1] Strachnov V., Valkovic V., Zeisler R. and Dekner R., Report on the Intercomparison Run IAEA-313: ^{226}Ra , Th and U in Stream Sediment. IAEA/AL/037, IAEA, Vienna, Austria 1991.

Issued & supplied by

Analytical Quality Control Services (AQCS)
Agency's Laboratories, Seibersdorf
International Atomic Energy Agency
P. O. Box 100
A-1400 Vienna, Austria

Prepared by

Z. Radecki, M. Campbell, K. I. Burns