


10 padomi aizsardzībā pret jonizējošo starojumu bērniem invazīvajā radioloģijā

1. Atcerieties: augoša bērna audi pret jonizējošo starojumu ir jutīgāki nekā pieaugušajiem.

Salīdzinot ar pieaugušajiem, laiks, kurā var izpausties apstarošanas sekas, bērniem ir ilgāks.



2. Pirms procedūras aprunājieties ar bērna vecākiem:

- pajautājiēt par iepriekš veiktajiem radioloģiskiem izmeklējumiem;
- atbildiet uz viņu jautājumiem par radiācijas drošību.

Patient's Name _____ MR# _____ Date of exam _____

Step Lightly Checklist

Review steps below before starting the procedure.


Safety is a team effort: don't be afraid to ask the necessary questions to ensure you are working as a team to keep radiation dose to patients and staff as low as possible.

Reducing radiation dose must be balanced with safe, accurate and effective completion of the procedure. Not all the steps below may be possible in each case, depending on patient size, technical challenge and critical nature of the procedure. Overall patient safety is most important. The goal is to minimize the dose to the patient while providing important and necessary medical care.

- ☐ Ask patient or family about previous radiation (record card downloadable at this link). Answer questions about radiation safety (patient consent brochure downloadable here).
- ☐ Use ultrasound when possible.
- ☐ Position hanging table shields and overhead lead shields prior to procedure with reminders during the case as needed.
- ☐ Operators and personnel wear well-fitted lead aprons, thyroid shield and leaded eye wear.
- ☐ Use pulse rather than continuous fluoroscopy when possible, and with as low a pulse as possible.
- ☐ Position and collimate with fluoroscopy off, tapping on the pedal to check position.
- ☐ Collimate tightly. Exclude eyes, thyroid, breast, gonads when possible.
- ☐ Operator and personnel hands out of beam.
- ☐ Step lightly: tap on pedal and review anatomy on last image hold rather than with live fluoroscopy when possible; minimize live fluoroscopy time.
- ☐ Minimize use of electronic magnification; use digital zoom whenever possible.
- ☐ Acknowledge fluoroscopy timing alerts during procedure.
- ☐ Use last image hold whenever possible instead of exposures.
- ☐ Adjust acquisition parameters to achieve lowest dose necessary to accomplish procedure: use lowest dose protocol possible for patient size, lower frame rate, minimize magnification, reduce length of run.
- ☐ Plan and communicate number and timing of acquisitions, contrast parameters, patient positioning and suspension of respiration with radiology and sedation team in advance to minimize improper or unintended runs.
- ☐ Move table away from X-ray tube in both planes. Move patient as close to detector in both planes.
- ☐ Use a power injector, or extension tubing if injected by hand.
- ☐ Move personnel away from table or behind protective shields during acquisitions.
- ☐ Minimize overlap of fields on subsequent acquisitions.
- ☐ After procedure: record and review dose.

3. Lai vērstu personāla uzmanību uz aizsardzību pret jonizējošo starojumu, pirms katras procedūras sastādiet pārbaudāmo lietu sarakstu.

4. Savlaicīgi izplānojiēt procedūras un darbību secību, lai novērstu neprecīzu vai pārtrauktu attēlu sēriju uzņemšanu un nevajadzīgu atkārtotu apstarošanu.



5. Iespēju robežās aizsargājiēt pacienta vairogdziedzeri, krūts dziedzerus, acis un gonādas





RPOP Posters webpage!

<https://rpop.iaea.org/RPOP/RPoP/Content/AdditionalResources/Posters/index.htm>

10 padomi aizsardzībā pret jonizējošo starojumu bērniem invazīvajā radioloģijā



6. Lietojiet optimālu tehniku:

- Pazeminiet kadru nomaiņas ātrumu. Ja iespējams, to samaziniet no 7,5 līdz 3 kadriem sekundē.
- Nelietojiet režģi bērniem, kuri vieglāki par 20 kg, ja iespējams.
- Samaziniet eksponēcijas laiku.
- Samaziniet lauku pārklāšanos atkārtotos izmeklējumos.
- Lietojiet šaurāku kolimāciju.
- Izvairieties no elektro-optiskā palielinājuma lietošanas




7. Papildu eksponēciju vietā, kad vien iespējams, izmantojiet pēdējā attēla saglabāšanas opciju (LIH-Last Image Hold)


8. Palieliniet attālumu starp pacientu un rentgena spuldzi un samaziniet attālumu starp pacientu un detektoru.



9. Lietojiet starojuma dozas reģistrācijas un dozas samazināšanas tehnoloģijas



10. Pēc procedūras novērtējiet un reģistrējiet starojuma dozu.



RPOP Posters webpage!

<https://rpop.iaea.org/RPOP/RPoP/Content/AdditionalResources/Posters/index.htm>