

L15b Validation of external personal monitoring with TLDs /OSL Systems



IAEA

International Atomic Energy Agency

How to validate

- Scientific approach
 - Systematic assessment of the influence of representative parameters of the result and uncertainty of a measurement method.
 - The selection of assessed parameters shall be based on published scientific and technical literature.
 - It may be supported by special investigations.

How to validate

- Comparative approach
 - Comparison with other, validated, methods.
 - Use of extended calibration routines with certified standard materials.
 - Participation in intercomparison exercises.
 - Simulation or modelling including expert judgement.

Validation planning

- Validation of external monitoring, using film batches, TLD, or similar devices can be done by:
 - Calibration of the readers and dosimeters by irradiating at a Secondary Standard Dosimetry Laboratory
 - Systematic investigations by using control dosimeters running with the field dosimeters whenever applicable
 - Laboratory intercomparison exercises
 - Performance Testing

Validation planning

- Start with method selection
- Define acceptable performance criteria
- Plan irradiations to cover the performance criteria
- Do the irradiations
- Evaluate the dosimeters
- Assess the performance criteria
- Issue validation statement

Performance criteria

- The performance criteria are recommended by
- ISO/IEC61066
- “Thermo luminescence dosimetry systems for personal and environmental monitoring” and
- IAEA Safety Guide No RS-G-1.3.
“Assessment of occupational exposure due to external sources of radiation”

Performance Requirements

- **Batch homogeneity**
- **Reproducibility**
- **Linearity**
- **Detection threshold**
- **Self Irradiation**
- **Residual signal**
- **Energy response**
- **Isotropy**

Further test, additionally to performance requirements:

Pre-irradiation fading


Post-irradiation fading

High dose linearity

Reader stability

Batch homogeneity

- The evaluated value for any one dosimeter in a batch shall not differ from the evaluated value for any other dosimeter in the batch by more than 30% for a dose equal to 10 times the required detection threshold limit.



Batch homogeneity


- **Method:**
 - TLD cards were cleaned with Methanol p.a. and left for drying for at least 2 hours
 - The annealed cards were exposed to Sr-90 (irradiator integrated in 6600 TLD Reader) and read out.
 - The value E was assessed for each dosimeter.
 - E_{\max} and E_{\min} were determined.

Reproducibility

- The coefficient of variation of the evaluated value shall not exceed 7.5% for each dosimeter separately and all n dosimeters collectively for a dose of 10 mSv.
- **Method:**
- Five dosimeters were prepared, irradiated to 10 mSv and read out. This procedure was repeated 10 times.

Linearity

- The response shall not vary by more than 10% over the range:
- $7 \text{ mg} \cdot \text{cm}^{-2}$: 0.5 mSv – 1 Sv (50 mrem – 100 rem)
- $100 \text{ mg} \cdot \text{cm}^{-2}$: 0.1 mSv – 1 Sv (10 mrem – 100 rem)



Linearity

Method

- Prepare, irradiate and read out four groups of dosimeters.
- The dosimeters were annealed before irradiation
- The value E was determined for each group of irradiated dosimeters
- The mean of the evaluated value E was calculated for each group.
- The standard deviation for all groups of dosimeters was calculated.

Detection threshold

- The detection threshold shall not exceed:
- P_e (7 mg · cm⁻²) 0.5 mSv
- P_e (1 000 mg · cm⁻²) 0.1 mSv

Method

- 30 dosimeters were prepared and read out.
- The value E was determined for each unirradiated dosimeter
- The mean of the evaluated value E was calculated
- The standard deviation for all 30 dosimeters was calculated

Self Irradiation

- After a storage period of 30 d, the zero point shall not exceed:
 - P_e (7 mg · cm⁻²) 0.5 mSv
 - P_e (1 000 mg · cm⁻²) 0.1 mSv

Self Irradiation

- Method
 - Prepare 30-40 dosimeters and store them for 30 days under standard test conditions in a location where the background dose rate is known.
 - Read out the dosimeters and determine the value E .
 - Calculate E and s_E .
 - Determine the conventional true value C_b (background irradiation)

Residual signal

- After irradiation with a conventional true value of 100 mSv, the required detection threshold shall not be exceeded and the response shall not change by more than 10% at a dose level of 2 mSv

Residual signal


- **Method**
 - Four dosimeters were prepared, irradiated to 100 mSv and read out.
 - The same dosimeters were irradiated to 2 mSv and read out.
 - The value E was determined for each dosimeter.
 - The mean value E and the standard deviation were calculated

Energy response

- When irradiated with photons in the range: 15 keV – 3.0 MeV, the evaluated value shall not differ from the conventional true value by more than 30%.
- When irradiated with beta rays in the range (E_{\max}) 0.5 MeV to 3.0 MeV, the response shall not vary by more than 30%.


Energy response

- Method
- Several dosimeters irradiated at the IAEA SSDL with the following radiation qualities:
N-40, N-80, N-120, N-150, S-Cs, S-Co



Isotropy (Directional dependence)


- When irradiated in two perpendicular planes with photons of (60 ± 5) keV, the mean value of the responses at an angle of incidence of 20° , 40° and 60° from normal shall not differ from the corresponding response for normal incidence by more than 15%




Isotropy (Directional dependence)

- **Method:**
 - Radiation quality N-15 to N-300, S-Cs, S-Co
 - Dose range of 1-5 mSv and 5-10 mSv and
 - Angles of incidence:
-30°, -45°, -60°, -70°, 0°, 30°, 45°, 60°, 70°

N-030 - Response at -70°, -45°, 0°, 30°, 45°, 60° and 70°






N-150 - Response at -70°, -45°, 0°, 30°, 45°, 60° and 70°



Reader stability

- The evaluated values of dosimeters read out 24 h apart and 168 h apart shall not differ from each other by more than 5% and 10% respectively.
- Test Method:
- The reader stability is checked by quality control of the PMT noise, the reference light and the RCF behavior over the time.

Reader stability



Resume

- In this lecture we have discussed on how to:
 - Perform validation of external personal monitoring with TLDs /OSL Systems using published scientific and technical literature based parameters